DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

NOV 2 0 2001

Soflex-Israelens Contact Lens Ltd.
c/o Kevin Walls, RAC
Regulatory Insight, Inc.
13 Red Fox Lane
Littleton, CO 80127

Re: K013469
Trade/Device Name: Eye-Q (xylofilcon A) Soft (hydrophilic) Multifocal Contact Lens for Daily Wear.
Regulation Number: 21 CFR 886.5925
Regulation Name: Soft (hydrophilic) Daily Wear Contact Lens
Regulatory Class: Class II
Product Code: LPL
Dated: October 16, 2001
Received: October 18, 2001

Dear Mr. Walls:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4613. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html

Sincerely yours,

A. Ralph Rosenthal

A. Ralph Rosenthal, M.D.
Director
Division of Ophthalmic and Ear,
Nose and Throat Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

ıown): K013469

Soflex Eye-Q (Xylofilcon A) Soft (hydrophilic) Multifocal Contact Lens for Daily Wear

The Eye-Q (Xylofilcon A) Soft (hydrophilic) Multifocal Contact Lens for Daily Wear is indicated for daily wear for the correction of refractive ametropia (myopia and hyperopia) and presbyopia in aphakic and not-aphakic persons with non-diseased eyes that may exhibit refractive an/or corneal astigmatism up to 2.00 diopters. The lens may be disinfected using a chemical disinfection system.

TE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED

ıcurrence of CDRH, Office of Device Evaluation (ODE)

Daniel W. C. Brown, Ph.D.

(Division Sign-Off)
Division of Ophthalmic Devices

510(k) Number K013469

CB